2-1" Spuds in top of Chamber
1-1" Spud-bottom of Chamber
2-1" Spuds-top of Chamber
1-1" Spud-bottom of Chamber
place on opp. side
11"
3'-0"
3'-4"
9"φ
8-1/4"x3/4" F.H. Machine scrs. on 7" B.C.
5"φ
6" (Fit to Glass)
7 1/2"φ
1" Thick x 6"φ Tempered Glass
Sealing Compound
3/32" Rubber Gasket
Detail Observation Window
5 Req. Scale 6"=1'-0"
Detail Of Door handles
1-Req. I.S. & O.S. Each Door Scale 3"=1'-0"
3/8" Pl.
9/16" Pl. Hd.
Typical Weld details
3"=1'-0"
15'-8"
1'-6 1/2"
6'-9 1/4" Pl.
1'-8 1/2" Pl.
3'-10" Pl.
1'-4"
4" S.F.
1" Spud
7'-0" O.D. Shell
84" D.R.
9/16"
(All heads Same)
1/8" I.P.S. Spuds Far Side
3 1/2" x 1/4" Bar-9" long Reif Pl. with 2-1" Pipe Nipples 5" long - Thded Both Ends 7-Pairs Req.-Spaced as Shown.
1"x3" Maple Strips D4S
2"x 1/4" Bars Notched 12" Cts For Drainage
4'-0"
4'-4"
18"R.
11" on O.S. Shell
2'-4"
MK. | No. PCS. | SECTION | LENGTH | CUT FROM
2 Coats Lead Chr. I.S. & O.S.
2 Coats Gray I.S. & O.S.
No. REQD. One
TEST LD-1,000,124
THIS DRAWING AND THE INFORMATION THEREON IS THE PROPERTY OF
GARY STEEL PRODUCTS CORPORATION
NORFOLK, VIRGINIA
CUSTOMER N.A.C.A.
ADDRESS Langley Field Va.
TITLE Medical Decompression Chamber
REVISIONS D C B A DATE
DRAWN: W.D.J.
DATE 11/1/43
CKD.
APP.
CHF. ENG.
SALES
SCALE 1"=1'-0" & Noted
S.O. NO. 14305
D-
FILE

JAK45